

CERTIFICATE

AGROLAB GmbH

Dr.-Pauling-Str. 1, 84079 Bruckberg

AGROLAB transport box maintains the temperature of samples at +4°C (+/- 2°C) for at least 24 hours



Object under test: transport box for samples
Contents: refrigerated AGROLAB bottle set (8 bottles) with representative water samples (incl. frozen or unfrozen cold packs)
Test equipment: Calibrated climatic chamber and 3 calibrated temperature data loggers (2 inside the thermal box, 1 outside the thermal box; for monitoring the temperature of the climate chamber)
Test profile: in accordance with **Afnor 48 b / 48 d**
Test performance: Over a period of 48 hours, the climatic chamber was subjected to the temperatures that can be reasonably expected during transport. (1x summer, 1x winter). During this period, the temperatures inside and outside the box were logged and subsequently evaluated. The cooling box has successfully passed the test.

Freigegeben:

Erstellt:



ppa. Richard Röser



i. A. Mathias Schendzielorz