







# Bottlelist water and drinking water

AGROLAB Agrar und Umwelt GmbH, Dr.-Hell-Str. 6, 24107 Kiel





























Article no. name	material cap colour	volume	GHS-label	containing preservation filling instruction	No. Bottles - parameter
 A004 Neutral	PE, white red	500 ml			1 I <sup>-</sup> , Br <sup>-</sup> , BrO <sub>3</sub> <sup>-</sup> , Cl <sup>-</sup> , Extinction/color, F <sup>-</sup> , NO <sub>2</sub> <sup>-</sup> , NO <sub>3</sub> <sup>-</sup> , o-PO <sub>4</sub> <sup>3-</sup> , Silicate, SO <sub>4</sub> <sup>2-</sup> , TOC/DOC, turbidity <b>1 PMI</b> 1 PFC <b>1 Cr(VI)</b> <b>4 PFC (0,001µg/l)</b> <b>1 Glyphosate/AMPA</b> <b>1 EDTA/NTA</b> <b>1 Dry residue</b> 1 Glowbacteria 2 BOD <b>4</b> solid material, settable solids 1 daphnia test 1 fishegg test 1 green alga test 2 alpha-activity, total 2 beta-activity, total 4 waterlens test
 A400 Organics	glass, green black	500 ml		fill only 90%	1 EOX, odor <b>1</b> anionic, cationic and non ionic detergents 1 PAH 1 PCB 1 TPH (GC) <b>2</b> Phthalates 3 Pesticides / plant protection agents 1 Lipophilic substances <b>2</b> Oil and fat 1 Drugs / Sweeteners 2 Organotin
 A002 Mibio	PE, white blue	250 ml		Thiosulphate, don't rinse, fill completely – don't overrun	1 E.coli, coliforms, bacterial count 20°C and 36°C + <b>if necessary one more parameter</b> 1 For each additional parameter ( <b>Ps. Aeruginosa, Enterococcus, Clost. Perfringens</b> )
 A207 Mibio (pools)	PE, white blue	250 ml		Thiosulphate, don't rinse, fill completely – don't overrun	<b>1</b> Scooped sample see A002
 A001 Legionella	PE, white blue	125 ml		Thiosulphate fill completely – don't overrun	1 Legionella
 A203 CO <sub>2</sub>	PET, clear white	250 ml		bubblefree to the brim	1 Bc <b>1</b> Ac, HCO <sub>3</sub> <sup>-</sup> , CO <sub>2</sub> , CO <sub>3</sub> <sup>2-</sup> , pH, conductivity 4 Tritium
 A200 AOX	glass, brown green	250 ml		HNO <sub>3</sub> don't rinse	1 AOX <b>2</b> AOX (waste water)
 A208 COD, N, P	PE, white white	250 ml		H <sub>2</sub> SO <sub>4</sub> don't rinse	1 COD, TNb, TKN, P, NH <sub>4</sub> <b>1</b> TOC (waste water)
 A102 Metals	PE, white black	100 ml		HNO <sub>3</sub> don't rinse	1 B, metals, <b>P (ICP)</b> <b>1</b> Hg

For further information please ask your responsible CRM. All preservations according to EN ISO 5667-3 and additional local criteria

QMF\_507\_AAUK\_006\_03\_EN Bottlelist water and drinking water; Stand: 08.08.2019; Seite: 1 / 3

# Bottlelist water and drinking water

AGROLAB Agrar und Umwelt GmbH, Dr.-Hell-Str. 6, 24107 Kiel








	A103 VOC	glass, clear silver	20 ml		H <sub>2</sub> SO <sub>4</sub> don't rinse Fill all to the brim	2 VOC (BTX/LHKW for waste water) 2 Methane, Ethane, Ethene
	A101 VOC (pools)	glass, light brown silver	20 ml		Thiosulphate don't rinse Fill all to the brim	2 VOC, (BTX/LHKW for drinking and bathing water, surface water), THM
	Article no. name	material cap colour	volume	GHS- label	containing preservation filling instruction	No. Bottles - parameter
	A201 Rn	glass, brown black	250 ml		bubblefree to the brim	1 Radon
	A104 Chlorate, Chlorite	PP, clear white	30 ml		NaOH don't rinse	1 Chlorite, Chlorate
	A202 Sulfide	PE, white green	250 ml	 	ZnAc+NaOH don't rinse	1 Sulfide
	A105 Sulfite	PE, white white	60 ml	 	EDTA don't rinse	1 Sulfit
	A106 Phenolindex	glass, brown white	60 ml		H <sub>2</sub> SO <sub>4</sub> don't rinse	1 Phenolindex
	A401 Phenoles	glass, green white	500 ml		H <sub>2</sub> SO <sub>4</sub> don't rinse	1 Phenoles
	A107 Fe II, Mn II	glass, clear black	100 ml	 	HCl don't rinse bubblefree to the brim	1 Fe (II), 1 Fe (III) (photom.), 1 Mn (II)
	A108 Cyanide	PE, white green	60 ml		NaOH don't rinse	1 Cyanide
	A009 Chlorophyll	PE, black black	500 ml			2 Chlorophyll
	A109 Oxygen	PET, clear white	125 ml		bubblefree to the brim, add solution 1+2	1 Oxygen
	A900 O <sub>2</sub> fixing solutions	Glas, brown white- red	100 ml	  	MnCl <sub>2</sub> (Solution 1) alkaline Iodide-Azide- Solution (Solution 2)	Solutions for O <sub>2</sub> -fixation in A109 1 Solution 1 1 Solution 2

For further information please ask your responsible CRM. All preservations according to EN ISO 5667-3 and additional local criteria

QMF\_507\_AAUK\_006\_03\_EN Bottlelist water and drinking water; Stand: 08.08.2019; Seite: 2 / 3

# Bottlelist water and drinking water

AGROLAB Agrar und Umwelt GmbH, Dr.-Hell-Str. 6, 24107 Kiel

	A204 CO <sub>2</sub> marble test	PET, clear black	250 ml		Marble don't rinse	1 CO <sub>2</sub>
	A402 Metalle UBA	PE, weiß, blue	500 ml		HNO <sub>3</sub> don't rinse	2 Metals by UBA
	A210 Neutral filtered	PE, white cyan	250ml		Filtration on site	1 Waste water: NO <sub>3</sub> <sup>-</sup> , NO <sub>2</sub> <sup>-</sup> , o-PO <sub>4</sub> <sup>3-</sup>
	A213 NH <sub>4</sub> , DOC	PE, white red	250ml		H <sub>2</sub> SO <sub>4</sub> Don't rinse	1 Waste water: NH <sub>4</sub> <sup>+</sup> , DOC
	A800 Radio	PE, White White	5000ml			1 I-131, Ra-226, Ra-228, Cs-134, Cs-137, Co-60, Pb-210

For further information please ask your responsible CRM. All preservations according to EN ISO 5667-3 and additional local criteria

QMF\_507\_AAUK\_006\_03\_EN Bottlelist water and drinking water; Stand: 08.08.2019; Seite: 3 / 3